

Declaration of Conformity

According to EN ISO/IEC 17050-1:2010

Manufacturer Name: National Instruments
Address: 11500 North MoPac Expressway
Austin, Texas, 78759 USA
Web: www.ni.com/certification

We hereby declare under our sole responsibility that the following apparatus:

Product Description: Ch-Ch Isolated Digital Input, Mechanical Relay Output Board for PXI
Model Number(s): NI PXI-6521
Product Category: Electrical equipment for measurement, control, and laboratory use.

Complies with the essential requirements of the following applicable European Directives:


Electromagnetic Compatibility (EMC) Directive 2014/30/EU,
Low-Voltage (Safety) Directive 2014/35/EU,

Conformity is assessed in accordance to the following standards:

EMC: Emissions
EN 61326-1:2013 (IEC 61326-1:2012), Class A
EN 55011:2010 (IEC CISPR 11:2010), Group 1, Class A
Immunity
EN 61326-1:2013 (IEC 61326-1:2012), Basic
EN 61000-4-2:2009 (IEC 61000-4-2:2008)
EN 61000-4-3:2010 (IEC 61000-4-3:2010)
EN 61000-4-4:2010 (IEC 61000-4-4:2010)
EN 61000-4-5:2006 (IEC 61000-4-5:2005)
EN 61000-4-6:2009 (IEC 61000-4-6:2008)
EN 61000-4-8:2010 (IEC 61000-4-8:2009)
EN 61000-4-11:2004 (IEC 61000-4-11:2004)
Safety: EN 61010-1 (IEC 61010-1)

Supplementary Information:

1. This product meets the EMC requirements of the United States (FCC Part 15, Class A), Canada (ICES-001, Group 1, Class A), and Australia/New Zealand (AS/NZS CISPR 11, Group 1, Class A).
2. This product meets the Safety requirements of the United States (UL 61010-1) and Canada (CAN/CSA-C22.2 No. 61010-1).



May 17, 2016, Austin, Texas USA

Date and Place of Issue

John Revell, Section Manager of Product Compliance and Reliability